



It is recommended that a clear space of 300mm / 11.81in is allowed around the BioTector in all directions.
1500mm / 59.00in should be free in front of the BioTector so that its access doors can easily be opened.

REV.	DETAILS	DATE		PROJECT: BIOTECTOR	SCALE: N.T.S.
x2	Imperial dimensions added	08/05/18			DATE: 06/03/2015
x1	First Issue	08/06/15		DWG. TITLE: Access Requirements	DR.BY: MP
					CHKD.BY: SOM
				DWG. NO: 81104131.DWG	VERSION: x2